



## Certificate of Conformity No: 210200293-CM-004

Issued 31-03-2021

Replaces NA

Valid until 31-03-  
2022

Page 1 of 2

# CERTIFICATE

### **NEN Spec 1-2: 2020-11-09**

Community masks  
requirements for producers and importers with respect to COVID-19

With this certificate Kiwa confirms that the face mask type

***Pactics Facewear R2 – Business Pro***

***Filter Type D***

from

***Pactics Europe B.V.***

Complies with NEN Spec 1-2: 2020-11-09 Community masks requirements for producers and importers with respect to COVID-19

The applicable face masks can therefore, under the strict conditions as stated in the NEN Certification scheme NCS 1-2: 2020-12-15 and the Kiwa NCS 1-2 certification agreement, carry the NCS certification mark for Community masks as listed below for the period this certificate is valid.

Ron Scheepers  
Kiwa

*This certificate consists of 2 pages.  
Publication of this certificate is allowed.*

Kiwa Nederland B.V.  
Wilmersdorf 50  
Postbus 137  
7300 AC Apeldoorn  
The Netherlands  
Tel. +31 88 998 33 93  
info@kiwa.nl  
www.kiwa.nl



Annex I: Details of the Certified Community masks



## NEN Spec 1-2: 2020-11-09

### Annex I: Details of the Certified Community masks

**Certificate number** ..... : 210200293-CM-004

**Report number**..... : 210200293

Date of issue ..... : 31-03-2021

---

**Testing laboratory**..... : KIWA Nederland B.V.

Testing location/address ..... : Wilmersdorf 50, 7327 AC Apeldoorn, The Netherlands

Applicant's name ..... : Pactics Europe B.V.

Address ..... : Nieuwe Keizersgracht 4  
1018 DR Amsterdam

**Scope**..... : NEN Spec 1-2: 2020-11-09 Community masks requirements for producers and importers with respect to COVID-19

---

> **Test item description** ..... : Pactics Facewear R2 – Business Pro; Filter Type D

Manufacturer ..... : Pactics Cambodja, as registered in certification file

Trade Mark ..... : Pactics Facewear

Model/Type reference ..... : Pactics Facewear R2 – Business Pro; Filter Type D

Special facts..... : BFE > 93,5% ; Breathability < 32Pa/Cm<sup>2</sup>

Test item number(s) ..... : Filter type D

Date receipt of test item(s) ..: 12-02-2021

Date(s) of testing ..... : 17-02-2021 – 24-03-2021

---

**Remarks** ..... :

---

**Summary**..... : Complies with the requirements as far as identified in the corresponding test report.

---